

Corrigendum -1

No: ACTREC/PUR/21-22/T-273

02/05/2022

## MANUAL TENDER NOTICE

Refer Manual Tender Notice No ACTREC/PUR/21-22/T-273 for “Fully Automated Immunohematology Analyzer 1 No. (Patient Blood Grouping, Cross matching and antibody Screening) (Reagent Rental)” at ACTREC.

1. During the pre-bid meeting (held on 04/04/2022) following changes have been incorporated in technical specification:

TENDER SRL.NO.	EXISTING TECHNICAL SPECIFICATION	Refined TECHNICAL SPECIFICATION TO BE READ AS
11	The system should have the facility for auto reading, capturing and interpreting results using suitable device and password protection	Refined technical specification to be read as “ <b>The system should have the facility for auto reading, capturing and interpreting results using suitable device and password protection for all the spectrum of tests mentioned in S No 54</b> ”
18	The system should have onboard cooling facility for the reagents to maintain the stability of reagents	Refined technical specification to be read as “ <b>The company should provide an undertaking/certificate regarding the stability of reagents and consumables at room temperature with duration</b> ”.
24	The system should have a mechanism to identify hemolysed, lipaemic or icteric samples with indication of same to the users	Refined technical specification to be read as “ <b>The system should allow the user to check and validate the results after processing of haemolysed, lipaemic or icteric samples</b> ”.
57	Cost Per Test (CPT) is to be indicated for the optional tests: Donor Blood grouping (Forward and Reverse), ABO antibody Titre (IgG and IgM), Extended phenotyping for Kidd, Duffy and MNS, Daratumumab discrepancy workup ,Weak D testing, AIHA work Up: IgG antibody titre determination	In refined technical specification to be read as “ <b>Optional Test: Cost Per Test (CPT) is to be indicated for the optional tests: 1. Donor Blood grouping (Forward and Reverse), 2. ABO antibody Titre (IgG and IgM), 3. ABO grouping for New-born, 4. Partial D phenotyping, 5. Minor antigen phenotyping for Kell, Kidd, Duffy, MNS and P, 6. DCT IgG dilution, 7. DCT IgG1/IgG3 dilution</b> ”
60	For the period of staff training, all the necessary reagents should be supplied with the system for the standardization and calibration for all the test free of cost	Refined technical specification to be read as “ <b>All the necessary reagents should be supplied with the system for the period of staff training along with standardization and calibration for all the test free of cost: ABO and Rh Grouping-100 test, AHG crossmatching-100 test, Antibody screening- 50 test, Direct Coombs Test- 50 test, Monospecific Direct Coombs Test- 20 test; Weak D Testing-20 test</b> ”.
61	Incase if, the workload may increase/ decrease as per requirement of department , placement of additional equipment	Refined technical specification to be read as “ <b>In case of increase in workload as per the requirement of department, vendor shall place an additional equipment in proportion</b> ”

TATA MEMORIAL CENTRE  
 ADVANCED CENTRE FOR TREATMETN, RESEARCH AND EDUCATION IN CANCER (ACTREC)  
**PURCHASE DEPARTMENT**

	with increasing workload at no additional cost shall be the responsibility of the vendor	<b>to increase in workload with no additional cost to the institute”.</b>
65	Provide List of consumables with their prices in the Financial bid.to conduct the above mentioned tests in S No. 44, 45 & 46	Refined technical specification to be read as <b>“Provide List of consumables with their prices in the Financial bid.to conduct the above mentioned tests in S No. 54”</b>
82	The equipment should be entirely maintained by the company with periodic (One visit every six months) preventive maintenance, calibrations and break down repairs including spare parts. The company has to ensure uptime guarantee of 95% taking into consideration 313 working days in a year.	Refined technical specification to be read as <b>“The equipment should be entirely maintained by the company with periodic (One visit every six months) preventive maintenance, calibrations once yearly and no limits for breakdown calls. The company has to ensure uptime guarantee of 95% taking into consideration 313 working days in a year”.</b>
90	Certificate to be provided to the effect that shut down period of the machine must not exceed for more than 48 hours & back up equipment option in case of equipment breakdown	Refined technical specification to be read as <b>“Certificate to be provided to the effect that shut down period of the machine must not exceed for more than 48 hours”.</b>

Further please note that all other Terms & Conditions of the tender remains same as per the Tender Documents.

**PURCHASE OFFICER**