

**TATA MEMORIAL CENTRE**  
Advanced Centre for Treatment, Research and Education in Cancer

Technical Specification of: Theatre Suction Trolley Units  
Qty: 28 Nos.

**TECHNICAL REQUIREMENT:**

Name of the Manufacturer: \_\_\_\_\_

Name of the Vendor: \_\_\_\_\_

Sr. No.	Description of item	Compliance
	Model Name:	
	Make:	
01	Whole system should be non electric.	
02	Double jar Suction trolleys.	
03	Suitable NFPA standard Probe /Connector (using in ACTREC Solid Tumor OTs) should be supplied with the unit.	
04	Each Suction jars should be more than 2000ml capacity	
05	Safety jar with overflow safety device should be available	
06	Metal brackets should be available for fixing the Suction jars properly.	
07	Castor wheels should be available for easy movement of Suction trolleys	
08	Good quality Vacuum regulator should be available	
09	Vacuum ON/OFF switch should be available.	
10	Vacuum gauge should be available to show the suction pressure.	
11	Suction jars bottom level should be atleast 1 feet from the ground level	
12	Handle should be available for easy movement.	
13	Trolley construction should be rust free SS 304 material.	
14	<b>Warranty: One year</b>	
15	<b>Free Service: NA</b>	

*Jaiswal* 