

TECHNICAL COMPLIANCE FORM			
Ref: Tender No:			
TATA MEMORIAL CENTRE			
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER			
TECHNICAL SPECIFICATION OF :- - "Analytical Balance " Qty - 03 Nos.			
Intended Use: For using in Hot Lab			
A	Technical Specifications / Scope of supply	COMPLIANCE 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.	REMARK
B	Name of the Manufacturer :		
C	Country of Origin:		
D	Model Name / Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each):		
E	Year of Introduction Internationally		
F	Year of Introduction in India		
G	Please quote the latest state of the art model which meets the tender specifications. The offer should be complete in all respect for the intended use.		
H	Technical Specifications/Scope of supply		
I	"Analytical Balance" , complete as per below mentioned configuration and specifications:		
J	Features and specifications required:		
1	Stainless steel weighing plate		
2	Measuring range 0.1 mg-120 g		
3	Measuring range resolution: 0.001 mg		
4	Single cell weighing system, high-stability mode, tare function		
5	Weighing pan W × D : 8 cm × 8 cm		
6	storage temp. : Room Temp		
7	Performance parameter required as below: (like accuracy, sensitivity, sample volume requirement, resolution or such performance parameters to be highlighted)		
8	Standard accessories (Which comes along with the machine, pricing of which is part of the equipment)		
9	Periodic Calibration		
10	The Equipment/ System should comply to US FDA/ European CE/ BIS/ ICMED certification. Please attach copy of the certificate(s).		
11	User's list: A list of installations with the address and contact numbers to be provided. (User list should be for the quoted model.)		
12	Input power supply requirements: 240 V AC ± 10 %, 50 Hz, single phase/ or suitable DC power adaptor . Specify if any other power supply requirement is recommended.		
13	After Sales Service Support.		
14	1. Complete operational manuals should be provided.		
15	Installation,Commissioning and Training is included in the scope of supply. Bidder, Manufacturer and/or its authorized representative should undertake installation and commissioning of the equipment.		
16	Warranty and after sales support		

17	Warranty : 1 Year comprehensive-on site- factory warranty for the entire equipment system.		
19	Important terms to be noted by the bidders:		
20	<u>Read the above scope of supply carefully and quote accordingly. Incomplete and /or partially complete offers are liable to be rejected.</u>		
21	Manufacturer/bidder should clearly indicate typical life span / Life expectancy of the quoted equipment.		
22	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.		
23	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.		