

TECHNICAL COMPLIANCE FORM			
Ref: Tender No:			
TATA MEMORIAL CENTRE			
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER			
TECHNICAL SPECIFICATION OF :- - " <b>Magnetic Stirrer with Hot Plate-</b> " Qty - 04 Nos.			
Intended Use: For Nuclear medicine Radiopharmacy use			
A	Technical Specifications / Scope of supply	<b>COMPLIANCE</b> 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.	REMARKS
B	Name of the Manufacturer :		
C	Country of Origin:		
D	Model Name / Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each):		
E	Year of Introduction Internationally		
F	Year of Introduction in India		
G	Please quote the latest state of the art model which meets the tender specifications. The offer should be complete in all respect for the intended use.		
H	Technical Specifications/Scope of supply		
I	" Magnetic Stirrer With Hot Plate " , complete as per below mentioned configuration and specifications:		
J	Features and specifications required:		
1	hotplate with stirrer		
2	Max stirring speed: ≥1200rpm		
3	Max volume: ≥1L; stirrer capacity: ≥1 L		
4	2 × 25 mm PTFE coated stir bars included		
5	stirring range: 100-MAX rpm		
6	temperature range : 0-450 °C, with thermostat		
7	high torque, speed stability, digital rpm and temperature setting with display		
8	Standard accessories (Which comes along with the machine, pricing of which is part of the equipment)		
9	Wide neck borosil test tubes (Set of 100)		
10	Periodic Calibration		
11	Regulatory Approvals, If any. Details with copies of approvals.		
12	Safety requirements:It should follow International / national safety requirement. Please specify certification with certifying agency and country with copies of certificate.		
13	The Equipment/ System should comply to US FDA/ CE/ BIS/ ICMED certification. Please attach copy of the certificate(s).		
14	User's list: A list of installations with the address and contact numbers to be provided. (User list should be for the quoted model.)		
15	Input power supply requirements: 240 V AC ± 10 %, 50 Hz, single phase. Specify if any other power supply requirement is recommended.		
16	After Sales Service Support.		

17	1. Complete operational manuals should be provided. 2. Company should provide technical support for hardware and software, spares and consumables. 3. Company should provide required spares and consumables for at least 7 years even after discontinuation of the model. 4. Bidders who do not qualify this criteria shall not be considered. Bidder should provide address of the authorized service centre and details of number of personnel with training credentials.		
18	<b>Warranty and after sales support</b>		
19	<b>Warranty : 2 Years</b> comprehensive-on site- factory warranty for the entire equipment system.		
20	<b>AMC: AMC is freezed at 1.5% per annum on ex works cost (exclusive of applicable GST) for 5 years after warranty.</b>		
21	<b>Important terms to be noted by the bidders:</b>		
22	<u>Read the above scope of supply carefully and quote accordingly. Incomplete and /or partially complete offers are liable to be rejected.</u>		
23	<b>Manufacturer/bidder should clearly indicate typical life span / Life expectancy of the quoted equipment.</b>		
24	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.		
25	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied ".		
26	<b>Declaration by the bidder</b>		
27	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.		