| TECHNICAL COMPLIANCE FORM  |  |  |         |  |  |
|--|--|--|---------|--|--|
| Ref: Tender No:  |  |  |         |  |  |
|  | TATA MEMORIAL CEN  |  |         |  |  |
| ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER  TECHNICAL SPECIFICATION OF : " Finn Pipette Variable Volume with Pipette Carousel " Qty - 06 sets |  |  |         |  |  |
| TECHNIC  | ALSPECIFICATION OF : Finn Pipette Variable Vol   | ume with Pipette Carousei Qty - 06 sets                              |         |  |  |
| Intended   | Use: For Radiopharmaceuticals lab in Nuclear Medic   | ine  |         |  |  |
| Α  | Technical Specifications / Scope of supply   | COMPLIANCE   | REMARKS |  |  |
|  |  | 1) Mention "Complied /Not Complied"                                  |         |  |  |
|  |  | 2) Highlight any deviations.   |         |  |  |
|  |  | 3) Mention part number/Catalogue number of the relevant item quoted. |         |  |  |
|  |  | indifficer of the relevant item quoted.                              |         |  |  |
|  |  |  |         |  |  |
| В  | Name of the Manufacturer :   |  |         |  |  |
| С  | Country of Origin:   |  |         |  |  |
| D  | Model Name / Model No. of the Equipment (If the  |  |         |  |  |
|  | equipment is a combination of more than one  |  |         |  |  |
|  | equipment then list down the model no.   |  |         |  |  |
|  | pertaining to each):   |  |         |  |  |
| E  | Year of Introduction Internationally   |  |         |  |  |
| F  | Year of Introduction in India  |  |         |  |  |
| G  | Please quote the latest state of the art model   |  |         |  |  |
|  | which meets the tender specifications. The offer   |  |         |  |  |
|  | should be complete in all respect for the  |  |         |  |  |
|  | intended use.  |  |         |  |  |
|  | Tack wise I Constitue time I Constitue time I  |  |         |  |  |
| H  | Technical Specifications/Scope of supply   |  |         |  |  |
| '  | "Finn Pipette Variable Volume with Pipette Carousel ", complete as per below mentioned     |  |         |  |  |
|  | configuration and specifications:  |  |         |  |  |
| j  | Features and specifications required:  |  |         |  |  |
| ,  | Pipettor volume - 0.1–2.5 μL, 2–20 μL, 20–200 μL,  |  |         |  |  |
| 1  | 100–1,000 μL, 0.5-5ml  |  |         |  |  |
| 2  | Carousel stand-6 nos.  |  |         |  |  |
| 2  | 2 full boxes of Eppendorf pipette tips for each pipette                                    |  |         |  |  |
| 3  | volume   |  |         |  |  |
| 4  | Pipettes should be single-channel volume   |  |         |  |  |
| 5  | volume measuring range Ø.1-2.5 μL-6 No.  |  |         |  |  |
| 6  | volume measuring range 2-20 µL -6 no.  |  |         |  |  |
| 7  | volume measuring range 20-200 µL-6 No.,  |  |         |  |  |
| 8<br>9   | volume measuring range 100-1000 μL-6No.<br>volume measuring range 0.5-5ml-6 No.            |  |         |  |  |
| 9  | Performance parameter required as below: ( like  |  |         |  |  |
| 10   | accuracy, sensitivity, sample volume requirement,  |  |         |  |  |
|  | resolution or such performance parameters to be  |  |         |  |  |
|  | highlighted)   |  |         |  |  |
|  | Standard accessories (Which comes along with the   |  |         |  |  |
| 11   | machine, pricing of which is part of the equipment)  |  |         |  |  |
|  | ,  |  |         |  |  |
|  | The Equipment/ System should comply to US FDA/   |  |         |  |  |
| 12   | European CE/ BIS/ ICMED certification. Please  |  |         |  |  |
|  | attach copy of the certificate(s).   |  |         |  |  |
| 13   | User's list: A list of installations with the address                                      |  |         |  |  |
|  | and contact numbers to be provided. (User list   |  |         |  |  |
|  | should be for the quoted model.)   |  |         |  |  |
| 14   | After Sales Service Support.   |  |         |  |  |
| 15   | Warranty and after sales support   |  |         |  |  |
| 16   | Warranty: 2 Years comprehensive-on site- factory warranty for the entire equipment system. |  |         |  |  |
| 40   | Important terms to be noted by the bidders:  |  |         |  |  |
| 18   |  |  |         |  |  |
| 19   | Read the above scope of supply carefully and quote   |  |         |  |  |
|  | accordingly. Incomplete and /or partially complete   |  |         |  |  |
|  | offers are liable to be rejected.  |  |         |  |  |
|  | Manufacturer/bidder should clearly indicate  |  |         |  |  |
| 20   | typical life span / Life expectancy of the quoted  |  |         |  |  |
|  | equipment.   |  |         |  |  |
|  |  |  |         |  |  |

| 21 | Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.  |  |
|----|--|--|
| 22 | After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown within 10 days from the date of request.   |  |
| 23 | Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected. |  |