

TECHNICAL COMPLIANCE FORM		
Ref: Tender No:		
TATA MEMORIAL CENTRE		
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER		
TECHNICAL SPECIFICATION OF :- - "Direct Laryngoscope -Qty-1 No.. "		
Intended Use: Direct laryngoscope need for Head and Neck surgery		
Sr. No.	Name of the Vendor :	
A	Technical Specifications / Scope of supply	COMPLIANCE 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
B	Name of the Manufacturer :	
C	Country of Origin:	
D	Model Name / Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each)	
E	Year of Introduction Internationally	
F	Year of Introduction in India	
H	Technical Specifications/Scope of supply	
I	"Direct Laryngoscope " , complete as per below mentioned configuration and specifications:	
J	Features and specifications required:	
1	Operating Layngoscope for Adult ., length 17cm - 1No	
2	Operating Laryngoscope for Adult , Length 18CM (For anterior commissure) -1 No	
3	Fiber Optic light carrier for distal illumination, working length 14cm for use with laryngoscope -1 Nos	
4	Laryngeal Biopsy Forceps , Heavy model with Oval jaws , Size 3 x 4mm Single action jaws , with cleaning connector working length 23cm -1 No	
5	Grasping Forceps without ratchet serrated straight with cleaning connector staright working length 23- 2 Nos	
6	Adequate length of compatible suction tips - 2No should be provided	
7	Light source with Fiber optic cable	
8	Lamp type should be LED	
9	Average lamp life should be 25,000 Hrs or more	
10	Power Supply - 230V AC, 50Hz	
11	Light Source weight should be ≤ 5kg	
12	Type of protection against electric Shock -Class I	
13	It Should have provision to adjust the light intensity as per needs	
14	Fiber optic light cable should be 3.5mm in diameter and atleast 2m long	
15	Fiber optic cable should be heat resistant	
16	All the instruments should be from same manufacturer with European CE/USFDA/ICMED/ BIS Certification	
17	International Standards: The unit should comply with International standards and should have European CE/US FDA /ICMED /BIS Certification	
18	Local Service Support : Should have local office and service support /service engineer for attending the local breakdown calls	
19	Response Time : Should not be more than 12 hours from lodging a breakdown complaint on toll free or by email	

20	Safety requirements: It should follow International / national safety requirement. Please specify certification with certifying agency and country with copies of certificate.	
21	User's list: A list of installations with the address and contact numbers to be provided. (User list should be for the quoted model.)	
22	Input power supply requirements: 240 V AC \pm 10 %, 50 Hz, single phase. Specify if any other power supply requirement is recommended.	
23	Installation,Commissioning and Training is included in the scope of supply. Bidder, Manufacturer and/or its authorized representative should undertake installation and commissioning of the equipment.	
24	Warranty and after sales support	
25	Warranty: 2 years comprehensive-on site- Factory warranty-for entire equipment. All third party major items, accessories/attachment etc. integrated with the system should also cover 2 years -factory warranty from respective manufacturers.	
26	Free AMC - Labour only : Free AMC for 8 Years Post warranty	
27	Manufacturer/bidder should clearly indicate typical life span / Life expectancy of the quoted equipment.	-- years
28	Mention typical site requirements for the installation.	--
29	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.	Agree/Not Agree
30	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. within 10 days from the date of request. Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)	Agree/Not Agree
31	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.	Agree/Not Agree
32	Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.	Agree/Not Agree
34	Declaration by the bidder	

35	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.	Yes/No
----	--	--------