

TECHNICAL COMPLIANCE FORM		
Ref: Tender No:		
TATA MEMORIAL CENTRE		
ADVANCED CENTRE FOR TREATMENT, RESEARCH & EDUCATION IN CANCER		
TECHNICAL SPECIFICATION OF :- "Tissue floating waterbath -Qty-1 No.. "		
Intended Use: To dissolve the wax during the tissue processing in Surgical Pathology lab		
Sr. No.	Name of the Vendor :	
A	Technical Specifications / Scope of supply	COMPLIANCE 1) Mention "Complied /Not Complied" 2) Highlight any deviations. 3) Mention part number/Catalogue number of the relevant item quoted.
B	Name of the Manufacturer :	
C	Country of Origin:	
D	Model Name / Model No. of the Equipment (If the equipment is a combination of more than one equipment then list down the model no. pertaining to each)	
E	Year of Introduction Internationally	
F	Year of Introduction in India	
H	Technical Specifications/Scope of supply	
I	"Tissue Floating waterbath" , complete as per below mentioned configuration and specifications:	
J	Features and specifications required:	
1	Should have a rectangular design	
2	Should have a capacity of minimum 2 litres.	
3	Jet black inner surface of the chamber with removable glass tray and illumination light source which provides better contrast to identify sections and easy to clean surface.	
4	Electronic temperature control for exact temperature regulation .	
5	Temperature indication by digital display.	
6	Temperature range : +30 to + 80 deg C	
7	Temperature should be adjustable by membrane keypad which should be less sensitive to water spillage, paraffin contamination.	
8	The water bath is illuminated from the side and provides exceptional viewing of the tissue section(s)	
9	Preferably should be supplied with 2 nos. of glass trays as standard.	
10	Outer parts should be made with corrosion free material.	
11	Should be USFDA /European CE/ ICED/BIS marked.	

CM
 21/11

Co

Signature

12	Should perform calibration of the equipment yearly during warranty which will include spares replacement, validation, consumable spares, Preventive Maintenance, Breakdown calls. Testing & measuring equipments used should be traceable to SI units through National /international Standards (as per NABL norms).	
13	Please specify footprint size & its weight.	
14	IQ/OQ/PQ should be provided by the vendor.	
15	Local Service Support: Should have local office and service support/ service engineer for attending the local breakdown calls.	
16	Response Time: Should not be more than 12 hrs from lodging a breakdown complaint on toll free or by email.	
17	Performance parameter required as below: (like accuracy, sensitivity, sample volume requirement, resolution or such performance parameters to be highlighted)	
18	On-Line UPS/Voltage Stabilizer requirements	
19	Periodic Calibration	
20	Safety requirements: It should follow International / national safety requirement. Please specify certification with certifying agency and country with copies of certificate.	
21	User's list: A list of installations with the address and contact numbers to be provided. (User list should be for the quoted model.)	
22	Input power supply requirements: 240 V AC \pm 10 %, 50 Hz, single phase. Specify if any other power supply requirement is recommended.	
23	Installation, Commissioning, testing and Training	
24	Installation, Commissioning and Training is included in the scope of supply. Bidder, Manufacturer and/or its authorized representative should undertake installation and commissioning of the equipment.	
25	Warranty and after sales support	
26	Warranty: 2 years comprehensive-on site- Factory warranty- for entire equipment. All third party major items, accessories/attachment etc. integrated with the system should also cover 2 years -factory warranty from respective manufacturers.	
27	Free AMC: Free for 5 years post warranty.	
28	Important terms to be noted by the bidders:	Agree/Not Agree
29	<u>Read the above scope of supply carefully and quote accordingly.</u> <u>Incomplete and /or partially complete offers are liable to be rejected.</u>	
30	Manufacturer/bidder should clearly indicate typical life span / Life expectancy of the quoted equipment.	-- years
31	Mention typical site requirements for the installation.	--
32	Item/Equipment quoted should not be a second hand/refurbished/on verge of obsolescence/outdated model. Similarly, after award of contract, the item supplied should be a newly manufactured item and not the old item already manufactured and stocked for more than 1 month.	Agree/Not Agree

33	After opening of the Technical bid (Part-1), Physical demonstration of the quoted model may have to be shown / arranged by the bidder, if requested by the Institute. Physical demonstration may be shown at ACTREC within 10 days from the date of request . Physical Demonstration may be requested to confirm the availability of any or all technical features as mentioned/stated in the technical bid. Physical Demonstration will also be a part of technical evaluation process. If the bidder does not comply, such bids are liable to be disqualified. (Demonstration of quoted model is to be shown and not the demonstration of similar models with different technical specifications and features)	Agree/Not Agree
34	Past experience of the bidders in terms of quality of supplied equipments, after sales service and application support will be taken into consideration while technical evaluation.. Bidders who has unsatisfactory past experience in last 2-3 years, in terms of quality of supplied equipments, after sales service and application support, bids of such bidders may liable to be rejected.	Agree/Not Agree
35	Complete and detailed information should be provided in respect to each point specified in the specifications. <u>Technical bids that are incomplete in any respect are liable to be rejected.</u> Provide relevant supportive information, publications, catalogue, etc. Bidders providing misleading or wrong information are liable to be rejected. All technical claims should be printed in the technical brochure of the equipment.	Agree/Not Agree
36	If any contradictory statements /figures/information is observed in the compliance chart and in the technical bid, then the technical information mentioned in the product literature/brochure will be considered true and further evaluation will be done based on the information given in the product literature/brochure. Regarding the warranty and free AMC terms, information quoted in the technical bid will only be taken as valid terms.	Agree/Not Agree
37	Compliance column may be filled with relevant data, figures, range etc. as applicable. Do not just mention "YES / NO / Complied ".	Agree/Not Agree
38	Declaration by the bidder	
39	We have quoted for all the items meeting the description/scope of supply in the Financial bid as per prescribed format of the Tender documents and we agree that Partial/incomplete offers are liable for rejection.	Yes/No

10/11/22

